

 0124
 DIAGNOSTILINE TEST DOAC'ite TUVASTAMISEKS URINIINIST

 0001
 12

⇒

⇒

⇒

(EST)

Skemaatiline esitus DOAC Dipstick

The diagram illustrates the components of a DOAC Dipstick. It is divided into two main sections: 'Käsitsemise ala' (Handling area) and 'Ala test-patjadega' (Area with test pads). The 'Käsitsemise ala' is represented by a long, thin rectangle. The 'Ala test-patjadega' is represented by a vertical stack of four numbered boxes (1, 2, 3, 4) with corresponding descriptions to their right.

Ala test-patjadega	Testpadi 4 trombiini inhibiitori tuvastamiseks
3	Testpadi 3 factor Xa inhibiitori tuvastamiseks
2	Testpadi 2 uriini värvuse hindamiseks (ilma reagentideta)
1	Testpadi 1 kreatiiniini hindamiseks

Kui testpadjakesed 3 ja 4 on mõlemad positiivsed "pos", on test kehtetu, sest on väga ebatõenäoline, et patsienti ravitakse samal ajal mõlemat tüüpi DOAC'itega.

0124	CE mark – seade vastab IVDR (EL) 2017/746 määruse Euroopa nõuetele.	LOT	Partii number		Seade patsiendilähedaseks testimiseks
IVD	In vitro diagnostika	REF	Kataloogi number		Seade ei ole mõeldud enesetestimiseks
	Tootja		Säilitustemperatuur		Ühekordseks kasutamiseks
	Pakendi sisu on piisav n-arvu testide jaoks		Aegumiskuupäev		Vaata kasutusjuhendit

 DOASENSE™ GmbH, Waldhofer Str. 102, 69123 Heidelberg, Germany
 Email: info@doasense.de, www.doasense.de